REF DX.X-1P29-25 DX/X-1B29

Rapid Response® Multi-Drug Test Panel

(Urine)



DRUG SCREENING

The Rapid Response® Multi-Drug Test Panel is a rapid, qualitative drug screening solution that can simultaneously detect multiple drugs in urine samples. The test is available in a variety of drug parameter and cut-off combinations and can test up to 14 parameters in single dip-card. Rapid Response™ Multi-Drug Test Panel is suitable for use by healthcare professionals.

Product Details

Sample: Urine

Format: Panel

Time-to-Result: 5 minutes

Storage Condition: 36-86°F/2-30°C

Test Principle: Lateral Flow Assay

Detection Window: 1-3 Days before collection

Key Features

- Test up to 14 parameters simultaneously
- Control and Test Regions marked for easy interpretation
- Integrated adulteration testing options available
- Longer detection window than saliva tests

Easy to Use

1. Collect the urine sample.
2. Submerge the panel strips vertically in the urine sample up to the indicated mark on the strips for 10-15 seconds.
3. Remove the panel from the sample and place the panel on a flat surface
4. Wait 5 minutes, then interpret the results.

These instructions are for illustrative purposes only, please read the product insert provided with your kit carefully before use.



Tel: +1 905-944-9565
Email: sales@btnx.com
www.btnx.com

Rapid Response® Multi-Drug Test Panel





Available Drug Parameters

- 6-MAM
- Amphetamines
- Barbiturates
- Buprenorphine
- Benzodiazepines
- Cocaine
- Cotinine
- Ecstasy (MDMA)
- EDDP
- Ethyl Glucuronide
- Fentanyl
- Hydromorphone
- Ketamine
- K2
- LSD
- Marijuana
- Methamphetamines
- Morphine
- Opiates
- Oxycodone
- Phencyclidine
- Propoxyphene
- Tramadol
- And more

Available Adulterant Tests

- Creatinine
- Glutaraldehyde
- Nitrite
- Oxidants
- pH
- Specific Gravity

Popular Parameter Combinations

Each Rapid Response® Multi-Drug Test Panel product has a set combination of parameters represented by a drug parameter code. The first three letters of a drug parameter code refer to the parameter detected; the number that follows represents the Cut-off Level in ng/mL for that parameter. Below is a selection most popular Health Canada Authorized Panels:

REF No.	Product Details	Drug Parameters	Quantity
D6.21-1B29-25	**6 Parameter** Test Panel with Adulteration Tests	COC150 EDDP100 MOP300 OXY100 BZO300 FYL10 + Cre/pH	25 Panels per Kit
D7.9.9-1B29-25	**7 Parameter** Test Panel with Adulteration Tests	COC150 EDDP100 MOP300 OXY100 BZO300 FYL10 MET500 + Cre/pH	25 Panels per Kit
D8.15-1B29-25	**8 Parameter** Test Panel with Adulteration Tests	MOP300 OXY100 FYL10 EDDP100 BUP10 MET500 COC150 BZO300 + pH/Cre	25 Panels per Kit
D9.9.3-1B29-25	**9 Parameter** Test Panel with Adulteration Tests	COC150 MET1000 BZO300 EDDP100 MOP300 OXY100 HMO1000 FYL10 BUP10 + Cre/pH	25 Panels per Kit
D10.24.1-1B29-25	**10 Parameter** Test Panel with Adulteration Tests	EDDP100 BUP10 MOP300 HMO500 OXY100 FYL10 COC150 MET1000 AMP1000 BZO300 + pH/Cre	25 Panels per Kit
D11.1-1B29-25	**11 Parameter** Test Panel with Adulteration Tests	COC150 AMP1000 EDDP100 MTD300 THC50 MDMA500 MOP300 OXY100 BZO300 FYL10 BUP10 + Cre/pH	25 Panels per Kit

Customizable parameter and cut-off combinations are available for qualifying orders.
To learn more contact your local Sales Representative, call us at +1-888-339-9964, or email us at sales@btnx.com.

Tel: +1 905-944-9565
Email: sales@btnx.com
www.btnx.com