


## Oral Fluid Lollipop Drug Test Quick Reference Guide



Scan For More Resources  
[btnx.com/product/oral-fluid-lollipop-drug-test](https://btnx.com/product/oral-fluid-lollipop-drug-test)



- Each strip is a column labeled **S#** (e.g., S1, S2, etc.).
- Test lines appear in rows labeled **T#** (e.g., T1, T2, etc.).
- The **Control line** is located in the top row labeled **C**.

### Preparation


Allow the test, specimen, and/or controls to reach room temperature (15-30°C) prior to testing. Instruct the donor to not place anything in the mouth, including food, drink, gum or tobacco products for at least 10 minutes prior to collection.

1




Remove the test from the sealed pouch and use it within one hour of opening.

2




Remove the cap to uncover the oral fluid collector.

3



Place the collector in your mouth. Do not let saliva touch the plastic body of the test. Keep the device level and do not tilt it. Hold it in your mouth until saliva flows across the test strips and the control lines (C) appear.


4



Remove the test from your mouth and put the cap on. Place it on a level surface and start the timer. Read the results. See Results Interpretation.

← See **Results Interpretation** on the back of this card.

## Results Interpretation



### How To Read The Table:

- Find the drug name in the Table.  
a) Identify the associated Strip (S#).  
b) Identify the Test region (T#).
- Match the associated coordinates S#T# from the Table to find the drug's position in the Results Window.
- Identify the Control line (C) on the same Strip (S#) of the drug's position in the Results Window.

BTNX Inc.  
Tel: +1 888-339-9964  
Email: support@btxn.com  
www.btxn.com  
Mon.-Fri. 9:00 AM - 5:30 PM EST

### Alcohol (ALC) Strip:

Interpret ALC strip between **1-2 minutes** by comparing the reaction pad with the enclosed color chart.

Blue or green



POSITIVE



NEGATIVE



INVALID

**Drug Tests:** Read results at **10 minutes**. Do not read results after 20 minutes. Use the table on the device to locate each drug and its corresponding position in the results window.

### NEGATIVE:

The Control (C) line is present and a test line appears in the Test region (T) for the drug. **Even a very faint line is considered negative.** A negative result indicates the drug is not present or is below the detectable level.


### POSITIVE:

The Control (C) line is present and **NO TEST LINE** appears in the Test region (T) for the drug. A positive result indicates the drug is present above the detectable level.


### INVALID:

**NO CONTROL LINE** appears in the Control region (C). Do not interpret invalid test Strips (S#). Repeat with a new test if needed.

### Examples:




All drugs are **positive**.




The drugs of S1T1, S2T2, S4T3, and S5T1 are **positive**.

The rest of the drugs are **negative**.



All drugs are **negative**.



**Negative:** S1T1, S1 T2, S1T3, S2T3, S4T1, S4T3

**Positive:** S2T1, S2 T2, S4T2

**Invalid:** S3T1, S3T2, S3T3, S5T1, S5T2, S5T3.